



Attestation of Conformity

No. N8A 067049 0083 Rev. 00

Holder of Certificate: **Datron HandelsGES m.b.h.**
Industriestrasse 1
9601 Arnoldstein, Austria

Product: **Laboratory Equipment
(PHARMACY REFRIGERATOR)**

Model(s): **YC50, YC50G, YC60, YC60G, YC110, YC110G,
YC120, YC120G, YC280, YC280G, YC360,
YC360G, YC440, YC440G**

Parameters: Rated Voltage: 220-240V
Rated Frequency: 50Hz
Rated Current:
1.2A for YC50, YC50G, YC60, YC60G, YC110,
YC110G, YC120, YC120G
1.5A for YC280, YC280G
1.6A for YC360, YC360G, YC440, YC440G
Refrigerant R600a
PS (High Low): 1,45MPa/0,85MPa
Protection Class: Class I
Degree of Protection: IP X0

Tested according to: EN 61010-1:2010/A1:2019
EN 61010-2-011:2017

The Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014*35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification.
For details see: www.tuvsud.com/ps-cert

Test report no.: 64111210159601

Date, 2022-10-20

Sergio Claudio Rasi